



# PERIYAR UNIVERSITY

SALEM- 636011, TAMIL NADU

NAAC A Grade - State University - NIRF Rank 73, ATAL Rank 4

Ref. No. PU/R/PL&D3/Biochemistry/20149/2021

Date: 28.12.2021

## TENDER NOTIFICATION

**Sealed Tenders** are invited by the Registrar, Periyar University, Periyar Palkalai Nagar, Salem – 11 for the **Purchase of Fluorescence Inverted Microscope for TANSCHÉ Project in the Department of Biochemistry** from reputed firms.

The detailed tender documents with specifications terms and conditions etc., can be downloaded from the University website **www.periyaruniversity.ac.in**. The last date for receipt of the tender in this office is **12.01.2022 at 2.00 p.m.**

## TENDER SUMMARY

Tender Ref. Number	PU/R/PL&D3/Biochemistry/20149/2021 Date: 28.12.2021
Tender Document Cost (Rs.) (Downloadable One)	Rs. 3360/- [(Demand Draft drawn in any Nationalized bank in favour of the Registrar, Periyar University, payable at Salem) [(Separate DD)]
Amount of EMD (Rs.)	1% of the quoted value [(Demand Draft drawn in any nationalized bank in favour of the Registrar, Periyar University, payable at Salem) [(Separate DD)]
Tender Cover Submission	upto 2.00 p.m. on <u>12.01.2022</u>

**REGISTRAR (FAC)**



# PERIYAR UNIVERSITY

SALEM - 636011, TAMIL NADU, INDIA  
NAAC A Grade - State University - NIRF Rank 73, ATAL Rank 4

From

The Registrar (FAC)  
Periyar University  
Periyar Palkalai Nagar  
Salem - 636 011.

To

**No.PU/R/PL&D3/Biochemistry/20149/2021**

**Date:**

**NAME OF THE SUPPLY: Purchase of Fluorescence Inverted Microscope for TANSCHÉ Project in the Department of Biochemistry**

## **Tender Conditions**

- 01) Sealed tenders will be received by the Registrar upto **2.00 p.m. on 12.01.2022** for the **“Purchase of Fluorescence Inverted Microscope for TANSCHÉ Project in the Department of Biochemistry”** as per the specifications given in the tender schedules.
- 02) The tender shall be submitted in sealed cover superscribed as **“Tender for Purchase of Fluorescence Inverted Microscope for TANSCHÉ Project in the Department of Biochemistry”** for use of Periyar University due on **12.01.2022 at 2.00 p.m.**
- 03) If the last day fixed for the receipt of tender happens to be a holiday tenders will be received on the next working day upto 2.00 p.m.
- 04) Each tender shall accompany with an **Tender Document Cost of Rs. 3360/-**. The Tender Document Cost in the form of the Demand Draft drawn in any nationalized bank in favour of the **Registrar, Periyar University, payable at Salem.** (Separate DD)
- 05) Each tender shall accompany with an **EMD at 1% of the Instrument Value**. The EMD in the form of the Demand Draft drawn in any nationalized bank in favour of the **Registrar, Periyar University, payable at Salem.** Tenders without E.M.D. will be summarily rejected. E.M.D. will not carry any interest.
- 06) **Tenders received late will be returned to the tenderer unopened.**
- 07) The entries in the tender schedules shall be as for as possible without scoring and corrections and over writings and shall be legible. The unavoidable correction or scoring shall be attested by full signature of the tenderer. The tenderer should sign on each page of the tender document.
- 08) **In the tender schedule, the tenderer should quote his rate and tax for each item separately in figures and words in the corresponding column.**

- 09) If the rates quoted in the schedule differ in words and figures, the lowest quoted rate will be taken.
- 10) No revision of rates will be accepted. Rates quoted shall be firm.
- 11) The rates quoted shall be for delivery at University Campus inclusive of charges such as packing and forwarding. Discount if any should be indicated prominently. The tenderer is solely responsible till delivery in good condition.
- 12) **The tenderer should produce the materials as per the specifications given in the tender schedule.**
- 13) Tender shall be submitted only in this official form and the tenderer should sign on each page of the tenderer enclosed without any omission. Tenders with price variation will not be accepted.
- 14) The tender shall be valid for a period of One year from the date of opening. Tenderer should not withdraw his tender after the tenders are opened. In case the tender is withdrawn after opening, the E.M.D. will be forfeited.
- 15) The E.M.D. of the unsuccessful tenderers will be refunded on their request immediately after the disposal of tender by the competent authority. **(Provide Your Bank Name & Address, Account Number, IFSC & MICR Number)**
- 16) **The successful tenderer shall also furnish security deposit at 2% of the value of the order. The Security Deposit will be refunded after one year from the date of supply of materials.**
- 17) In case of failure by the tenderer to supply items demanded within the period prescribed, the Registrar shall have the power to purchase from others in lieu of rejected or not delivered goods, the excess cost and expenses will be worked out and recovered from the tenderer.
- 18) Successful tenderer should execute an agreement on stamp paper to the value of ₹ 20/- within seven days from the date of receipt of intimation about acceptance of the tender. Failure to execute the agreement in the stipulated time will entail in forfeiture of the E.M.D.
- 19) Any dispute arising out of this contract shall be settled only on the court having jurisdiction of Salem.
- 20) The authority competent to accept the tender reserves the right to reject the tender without assigning reasons thereof.
- 21) Regarding the acceptance of supply with reference to the specifications and quality of materials supplied, the decision of the Registrar shall be final. The rejected materials should be removed within 15 days at Tenderer's cost.
- 22) This University's General rules for the supply of the materials and works will apply on this purchase also.

- 23) The goods should be supplied within 15 days of the receipt of order.
- 24) If the tenderer fails in the due performance of his supply within the time fixed, the tenderer is liable to pay as liquidated damages upto 2% per month for the supply value of such portion of the materials as have not been delivered.
- 25) The tender schedule is not transferable and it should be used only by the tenderer to whom it is officially issued.
- 26) Percentage of payment to be withheld for the effective performance of the contract, provided that withheld amounts do not exceed ten percent of the total value of contract.
- 27) The cost must include the warranty maintenance for 5 years from the date of intallation.
- 28) Complete address of the firm including year of establishment, Phone No., Fax No., E-mail address etc., may be furnished.
- 29) The tenderers should be responsible for erection & installation of the equipment at destination site and for making it fully operational. Payment and terms and conditions if any for the same would be specified by the tenderer separately.
- 30) Canvassing in connection with tender/quotation is strictly prohibited.
- 31) List of details of works executed and list of customers shall be attached.
- 32) The Company TIN/GST/PAN/CST Numbers must be mentioned in your Quotation.
- 33) Provide Your Bank Name & Address, Account Number, IFSC & MICR Number for EMD & Security Deposit Return and Payment Purpose.

This tender is submitted subject to agreeing to the above conditions.

**Details of Tender Document Cost (DD) Enclosed**

S.No.	Demand Draft No. / Date & Name of the Bank	Amount

**Details of EMD Enclosed**

S.No.	Demand Draft No. / Date & Name of the Bank	Amount

## ***TENDER SCHEDULE***

### **Purchase of Fluorescence Inverted Microscope for TANSICHE Project in the Department of Biochemistry**

The quantities given below are approximate and are likely to be increased/reduced.  
The rates should be quoted for the brand mentioned only.

**The rates should be inclusive of all taxes. Taxes must be exhibited separately in the bill. If not, deduction of tax (GST) will be made at the final payment.**

The Unit rates noted below are those governing payments.

The rates quoted are for delivery at the University Campus

<b>Sl. No.</b>	<b>Description of the Equipment</b>	<b>Qty.</b>	<b>Rate in Rs. (Without Tax)</b>	<b>Amount in Rs. (With Tax)</b>
1.	Fluorescence Inverted Microscope  (Specifications enclosed as Annexure)	1 No.		

**Specification**  
**Fluorescence inverted microscope**

<b>1</b>	<b>Viewing Head</b>	45° inclined trinocular head, 360° rotatable; interpupillary distance 50-75mm; Splitting ratio for Binocular:Trinocular = 100:0 or 0:100
<b>2</b>	<b>Contrasting Method</b>	Microscope should be equipped for Bright Field, Phase Contrast & Fluorescence contrast.
<b>3</b>	<b>Eyepiece</b>	High point eyepiece wide-field plan eyepiece PL10X/22 mm, diopter adjustment
<b>4</b>	<b>Objectives</b>	Infinity long working distance plan fluorescent objective. LWDPL4X/0.13/WD18.52mm Infinity long working distance plan fluorescent objective. LWDPL10X/0.30/WD7.11 mm Long working distance infinity plan semi-apochromatic phase contrast fluorescent Objective. LWDPL20X/0.45WD5.60 mm Long working distance infinity plan semi-apochromatic phase contrast fluorescent Objective. LWDPL40X/0.65/WD1.61 mm
<b>5</b>	<b>Stand and Stage</b>	Reflection and Transmission Frame; Coaxial focus system with tension adjustment, quintuple nosepiece, coarse range: 9mm, coke-oriented 6.5mm, 2.5mm down, fine precision: 0.002 mm; fixed Stage size: 215 mm x 250 mm; Three-band Filter switching mechanism. External wide Voltage transformer, input 100-240V, output 12V5A;
<b>6</b>	<b>Illumination</b>	5W LED illumination [long life (approx. 40000 hrs or more)] and Intensity adjustable. Broad-spectrum for imaging most common fluorescent stains. Spectral coverage is from the UV to the Red region.
<b>7</b>	<b>Condenser</b>	N.A.0.3 condenser, working distance 72mm, dismountable Centre-adjustable phase contrast slide plate for N.A.0.3, with 2 handle Centering telescope $\phi$ 30, Glass plate, Metal plate, Mechanical Mobile ruler, moving range 120(X)*78(Y)mm
<b>8</b>	<b>Fluorescence Filters</b>	Pixel shift corrected fluorescence filters for DAPI Stain/Dye, FITC/GFP/Alexa488 and TRITC/RHOD/Cy3.
<b>9</b>	<b>Attachment</b>	Stage extension plate, 35 mm Petri dish holder, slides holder( $\phi$ 54 and 26.5x76.5), Teraseki holder( $\phi$ 65 and 56X81.5), LBD color temperature change filter ( $\phi$ 45mm, IF550 green interference filter ( $\phi$ 45mm)
<b>10</b>	<b>Camera attachment</b>	HDMI camera – HDMI output Microscope camera with mouse + SD card slot. Sensor size : Sony IMX920,1/2.8”(5.57x4.1 30U). Pixel ( $\mu$ m) : 2.9x2.9. G Sensitivity: 1300mv with 1/30 s/0.15 mv with 1/30s. FPS/Resolution: 26@1920*1080P(HDMI), 26@1920*1080(USB) Binning: 1x1, Exposure: 0.036ms-8s. Output : USB+HDMI+SD card

<b>11</b>	<b>Software</b>	Licensed imaging software.
<b>12</b>	<b>Warranty</b>	1 year
<b>13</b>	<b>Installation</b>	Installation shall be provided free of cost at our site. After successful completion of commissioning of the equipment, the supplier shall demonstrate the functionality of system.
<b>14</b>	<b>Training</b>	The supplier shall provide comprehensive free onsite training for system operation along with routine maintenance requirement.
<b>15</b>	<b>Accessories</b>	Dust Cover & all other cables and connectors should be supplied along with instrument for hassle free operation.